

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Analysis For Total Metals By EPA Method 200.8

Client ID: M117973
Date Received: 11/20/07
Date Extracted: 11/26/07
Date Analyzed: 11/26/07
Matrix: Water
Units: ug/L (ppb)

Client: Alaskan Copper Works
Project: PO M117973, F&BI 711285
Lab ID: 711285-01 x10
Data File: 711285-01 x10.026
Instrument: ICPMS1
Operator: HR

Internal Standard:	% Recovery:	Lower Limit:	Upper Limit:
Germanium	118	60	125

Analyte:	Concentration ug/L (ppb)
Chromium	377
Nickel	311
Copper	332
Zinc	<10

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Analysis For Total Metals By EPA Method 200.8

Client ID: Method Blank
Date Received: Not Applicable
Date Extracted: 11/26/07
Date Analyzed: 11/26/07
Matrix: Water
Units: ug/L (ppb)

Client: Alaskan Copper Works
Project: PO M117973, F&BI 711285
Lab ID: I7-438 mb
Data File: I7-438 mb.016
Instrument: ICPMS1
Operator: HR

Internal Standard:
Germanium

% Recovery:
104

Lower
Limit:
60

Upper
Limit:
125

Analyte:	Concentration ug/L (ppb)
Chromium	<1
Nickel	<1
Copper	<1
Zinc	<1

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Date of Report: 11/27/07

Date Received: 11/20/07

Project: PO M117973, F&BI 711285

QUALITY ASSURANCE RESULTS FOR THE ANALYSIS OF WATER SAMPLES FOR TOTAL METALS USING EPA METHOD 200.8

Laboratory Code: 711252-01 (Duplicate)

Analyte	Reporting Units	Sample Result	Duplicate Result	Relative Percent Difference	Acceptance Criteria
Chromium	ug/L (ppb)	<1	<1	nm	0-20
Nickel	ug/L (ppb)	<1	<1	nm	0-20
Copper	ug/L (ppb)	4.28	3.72	14	0-20
Zinc	ug/L (ppb)	21.8	21.6	1	0-20

Laboratory Code: 711252-01 (Matrix Spike)

Analyte	Reporting Units	Spike Level	Sample Result	Percent Recovery MS	Acceptance Criteria
Chromium	ug/L (ppb)	20	<1	89	50-150
Nickel	ug/L (ppb)	20	<1	94	50-150
Copper	ug/L (ppb)	20	4.28	88 b	50-150
Zinc	ug/L (ppb)	50	21.8	91 b	50-150

Laboratory Code: Laboratory Control Sample

Analyte	Reporting Units	Spike Level	Percent Recovery LCS	Acceptance Criteria
Chromium	ug/L (ppb)	20	98	70-130
Nickel	ug/L (ppb)	20	94	70-130
Copper	ug/L (ppb)	20	95	70-130
Zinc	ug/L (ppb)	50	91	70-130

Data Qualifiers & Definitions

a - The analyte was detected at a level less than five times the reporting limit. The RPD results may not provide reliable information on the variability of the analysis.

A1 - More than one compound of similar molecule structure was identified with equal probability.

b - The analyte was spiked at a level that was less than five times that present in the sample. Matrix spike recoveries may not be meaningful.

ca - The calibration results for this range fell outside of acceptance criteria. The value reported is an estimate.

c - The presence of the analyte indicated may be due to carryover from previous sample injections.

d - The sample was diluted. Detection limits may be raised due to dilution.

ds - The sample was diluted. Detection limits are raised due to dilution and surrogate recoveries may not be meaningful.

dv - Insufficient sample was available to achieve normal reporting limits and limits are raised accordingly.

fb - The analyte indicated was found in the method blank. The result should be considered an estimate.

fc - The compound is a common laboratory and field contaminant.

fp - Compounds in the sample matrix interfered with quantitation of the analyte. The reported concentration may be a false positive.

hr - The sample and duplicate were reextracted and reanalyzed. RPD results were still outside of control limits. The variability is attributed to sample inhomogeneity.

ht - The sample was extracted outside of holding time. Results should be considered estimates.

ip - Recovery fell outside of normal control limits. Compounds in the sample matrix interfered with the quantitation of the analyte.

j - The result is below normal reporting limits. The value reported is an estimate.

J - The internal standard associated with the analyte is out of control limits. The reported concentration is an estimate.

jl - The analyte result in the laboratory control sample is out of control limits. The reported concentration should be considered an estimate.

jr - The rpd result in laboratory control sample associated with the analyte is out of control limits. The reported concentration should be considered an estimate.

lc - The presence of the compound indicated is likely due to laboratory contamination.

L - The reported concentration was generated from a library search.

nm - The analyte was not detected in one or more of the duplicate analyses. Therefore, calculation of the RPD is not applicable.

pc - The sample was received in a container not approved by the method. The value reported should be considered an estimate.

pr - The sample was received with incorrect preservation. The value reported should be considered an estimate.

ve - The value reported exceeded the calibration range established for the analyte. The reported concentration should be considered an estimate.

vo - The value reported fell outside the control limits established for this analyte.

x - The pattern of peaks present is not indicative of diesel.

y - The pattern of peaks present is not indicative of motor oil.

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

James E. Bruya, Ph.D.
Charlene Morrow, M.S.
Yelena Aravkina, M.S.
Bradley T. Benson, B.S.
Kurt Johnson, B.S.

3012 16th Avenue West
Seattle, WA 98119-2029
TEL: (206) 285-8282
FAX: (206) 283-5044
e-mail: fbi@isomedia.com

November 27, 2007



INVOICE #07ACU1127-2

Accounts Payable
Alaskan Copper Works
628 South Hanford
Seattle, WA 98134

RE: Project PO M117973, F&BI 711285 - Results of testing requested by Gerry Thompson for material submitted on November 20, 2007.

1 sample analyzed for Total Chromium, Copper, Nickel and Zinc by Method 200.8 @ \$80 per sample	\$ 80.00
Rush Charges (4 day) 60% of \$80.00	<u>48.00</u>
Amount Due	\$ 128.00

FEDERAL TAX ID # (b) (6)

711285

SAMPLE CHAIN OF CUSTODY ME 11-20-07 AI4

Send Report To GERARD Thompson
 Company ALASKAN COPPER WORKS
 Address 628 S. HANFORD ST
 City, State, ZIP SEATTLE WA 98134
 Phone # 206-351-6033 Fax # 206-382-4209

SAMPLERS <u>(signature)</u>	
PROJECT NAME/NO. <u>M810 Self Monitor</u>	PO # <u>M-11773</u>
REMARKS	

Page # _____ of _____
TURNAROUND TIME <input type="checkbox"/> Standard (2 Weeks) <input checked="" type="checkbox"/> RUSH <u>40%</u> Rush charges authorized by: _____
SAMPLE DISPOSAL <input type="checkbox"/> Dispose after 30 days <input type="checkbox"/> Return samples <input type="checkbox"/> Will call with instructions

Sample ID	Lab ID	Date	Time	Sample Type	# of containers	ANALYSES REQUESTED										Notes
						TPH-Diesel	TPH-Gasoline	BTEX by 8021B	VOCs by 8260	SVOCs by 8270	HFS	OC	OC	OC	OC	
M11773	01	11/20/07	9:00	H2O	1											

Friedman & Bruya, Inc.
 3012 16th Avenue West
 Seattle, WA 98119-2029
 Ph. (206) 285-8282
 Fax (206) 283-5044

SIGNATURE	PRINT NAME	COMPANY	DATE	TIME
Relinquished by: <u>(signature)</u>	<u>G. Thompson</u>	<u>ACW</u>	<u>11/20/07</u>	<u>10:05</u>
Received by: <u>(signature)</u>	<u>Nhan Phan</u>	<u>Fe BT</u>	<u>11/20/07</u>	<u>10:05</u>
Relinquished by:				
Received by:				

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

James E. Bruya, Ph.D.
Charlene Morrow, M.S.
Yelena Aravkina, M.S.
Bradley T. Benson, B.S.
Kurt Johnson, B.S.

3012 16th Avenue West
Seattle, WA 98119-2029
TEL: (206) 285-8282
FAX: (206) 283-5044
e-mail: fbi@isomedia.com

November 27, 2007

Gerry Thompson, Project Manager
Alaskan Copper Works
628 South Hanford
Seattle, WA 98134

Dear Mr. Thompson:

Included are the results from the testing of material submitted on November 20, 2007 from the Metro Self Monitor, PO M117973, F&BI 711285 project. There are 4 pages included in this report. Any samples that may remain are currently scheduled for disposal in 30 days. If you would like us to return your samples or arrange for long term storage at our offices, please contact us as soon as possible.

We appreciate this opportunity to be of service to you and hope you will call if you should have any questions.

Sincerely,

FRIEDMAN & BRUYA, INC.



Michael Erdahl
Project Manager

Enclosures
ACU1127R.DOC